



Certificate of Traceability

Level I Laboratory Certification

This certifies that ACM Global Laboratories, using Tosoh G8 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: G8	Calibrator Lot: ZS9001, ZS9002	Column Lot: EJ6367N
Reagent Lot: N8-160E, N8-258E, N8-357E, HW-636-NU	Calibrator Assigned Values: 5.77%, 10.46%	

Date of Certification: March 1, 2021 Certification Expires: March 1, 2022

Daniel Sault *Randie R Little PhD.* *Shawn Ansell*
NGSP Steering Committee Chair NGSP Network Coordinator SRL director/ supervisor